

Declaration of Conformity



| | | |
|------------------------------------|---|---------------------|
| Manufacturer Name | SD Biosensor, Inc. | |
| Manufacturer Address | <u>Head Office</u> C-4th&5th, 16, Deogyong-daero, 1556beon-gil, Yeongtong-gu, Suwon-si, Gyeonggi-do 16690, KOREA | |
| | <u>Manufacturing Site</u> 74, Osongsaengmyeong 4-ro, Osong-eup, Heungdeok-gu, Cheongju-si, Chungcheongbuk-do 28161, KOREA | |
| EC Representative Name | MT Promedt Consulting GmbH | |
| EC Representative Address | Altenhofstrasse 80 66386 St. Ingbert Germany | |
| Common Name | Rapid Test Kit | |
| Product Name | SARS-CoV-2 Rapid Antibody Test <i>*Please refer to "Annex I. Product List" on page 2 in more detail.</i> | |
| Reference Number | 9901-NCOV-02C | |
| Classification | Others not covered by Annex II and self-testing according to Directive 98/79/EC | |
| Conformity Assessment Route | Annex III of Directive 98/79/EC (EC Declaration of Conformity) | |
| Applied Standards | EN ISO 13485:2016 | EN ISO 18113-1:2011 |
| | EN ISO 14971:2012 | EN ISO 18113-2:2011 |
| | EN ISO 23640:2015 | EN ISO 15223-1:2016 |
| | EN ISO 17511:2003 | EN 62366:2008 |
| | EN 13612:2002 | |

We herewith declare that the above mentioned products meet the provisions of the Council Directive 98/79/EC for in vitro diagnostic medical devices. All supporting documentations are retained under the premises of the manufacturer. SD Biosensor, Inc. is exclusively responsible for the declaration of conformity.

Place: Suwon-si, Republic of Korea
Valid from: July 03, 2020

Signature

Hyo-Keun, Lee
CEO / President

Annex I. Product List

9901-NCOV-02C

SARS-CoV-2 Rapid Antibody Test

- Test Device (individually in a foil pouch with desiccant)
- Buffer bottle
- Capillary tube(20µl)
- Film
- Quick Reference Guide
- Instructions for use

EDMA Code

15 70 90 90 00

Description of EDMA code

Other Other Virology Rapid Tests